

HEALTH ACCURATE DIAGNOSTICS

Calibration Record

Format No: HAD-LAB/FM/44

Sl. No	Discipline	Name of equipment	UID (unique identificati on number)	Sr. No	Model/M ake	year of make	Date of Installation	Range and Accuracy	Date of last calibration	Calibration Due on	Calibrated By
	<i>Handwritten</i>										
1.	Clinical Biochemistry	Biochemistry Analyzer	HAD/C B/13	27002 550	Vitros 250/Ortho clinical diagnostics	2021	03-04-2021	As defined in scope	05/04/2022	05/10/2022	Ortho clinical diagnostics
2.	Clinical Biochemistry	Nephelometry Analyzer	HAD/C B/14	21191 64385 8	Mispa i2/Agappe	2019	18-07-2020	As defined in scope	18.01.2022	17.07.2022	Agappe Supplied
3.	Clinical Biochemistry	Electrolyte Analyzer	HAD/C B/15		Microlyte	2022	17/03/2022	As defined in scope	Self-calibration is done.		
4.	Clinical Biochemistry	Biochemistry Analyzer	HAD/C B/1	83105 C032 7	A15 Biosystem	2019	25-03-2019	As defined in scope	20.07.2021	19.07.2022	Biosystems
5.	Clinical Biochemistry	HPLC Analyzer	HAD/H /13	DJ4C 01202 1	Biorad D10/Biorad	2014	14-10-2014	As defined in scope	21/07/21	20/07/2022	Anand Chemiceutics
6.	Clinical Biochemistry	HPLC Analyzer	HAD/H /14	DM21 C290 01	Biorad D10/Biorad	2021	11.10.2021	As defined in scope	11.10.2021	11.10.2022	Anand Chemiceutics
7.	Clinical Biochemistry	Immunoassay analyzer	HAD/I A/6	iSR60 760	Abbott Architect	2017	10-08-2017	As defined in scope	04.01.2022	04.07.2022	Abbott

Handwritten
R. PARAG BINIWALE
M.D., (OB-GYN)
Reg. No. 62623

Issue No. 02	Issue Date: 05.07.2021	Prepared By: <i>Handwritten</i> QM	Copy No.	Page 1 of 23
Rev. No.: 00	Rev. Date: Nil	Approved By: <i>Handwritten</i>	Issued By: <i>Handwritten</i>	

Handwritten Signature
- Director,
PATANKAR HOSPITAL PVT. LTD.
PUNE



HEALTH ACCURATE DIAGNOSTICS

Calibration Record

Format No: HAD-LAB/FM/44

Sl. No	Discipline	Name of equipment	UID (unique identificati on number)	Sr. No	Model/M ake	year of make	Date of Installation	Range and Accuracy	Date of last calibration	Calibration Due on	Calibrated By
					I1000SR/ Abbott Architect						
8.	Clinical Biochemistry	Immunoassay analyzer	HAD/I A/2	600007 39	Delfia Express/ Perkin Elmer	2020	17-03-2020	As defined in scope	23.12.2021	22.12.2022	Perkin Elmer
9.	Serology	ELISA Plate Reader	HAD/S ERO/0 1	332- 1141	STAT FAX 332 PLUS			As defined in scope	Not Applicable		
10.	Serology	ELISA Plate Reader	HAD/S ERO/0 2	SN 452212 034	LISA Plus Micropla te Reader	2022	01/04/2022	As defined in scope	Not Applicable		

Patankar

**Director,
PATANKAR HOSPITAL PVT. LTD.**

Parag

DR. PARAG BINIWALE
M.D., (OB-GYN)
Reg. No. 62623

Issue No. 02	Issue Date: 05.07.2021	Prepared By: <i>Mangesh</i> QM	Copy No.	Page 2 of 23
Rev. No.: 00	Rev. Date: Nil	Approved By: <i>Mangesh</i> : <i>Mangesh</i>	Issued By: <i>Mangesh</i> (<i>Mangesh</i>)	

HEALTH ACCURATE DIAGNOSTICS

Calibration Record

Format No: HAD-LAB/FM/44

Sl. No	Discipline	Name of equipment	UID (unique identification number)	Sr. No	Model/Make	year of make	Date of Installation	Range and Accuracy	Date of last calibration	Calibration Due on	Calibrated By
11.	Clinical Biochemistry	Micropipette	HAD/IP/34	QW025 96	Agappe (Mispa-I Series)/ Agappe	2021	15-07-2021	2-20 µL	15.07.2021	15.07.2022	Eppendorf
12.	Clinical Biochemistry	Micropipette	HAD/IP/36	YN191 AE000 4150	Agappe (Mispa-I Series)/ Agappe	2021	15-07-2021	20-200 µL	15.07.2021	15.07.2022	Eppendorf
13.	Clinical Pathology	Centrifuge	HAD/C P/2	ZGAN- 03135	R8C Plus/ REMI	2020	27-05-2020	100- 4500 RPM	20.07.2021	19.07.2022	Ultra Air Solutions
14.	Accession	<u>Centrifuge</u>	<u>HAD/ MB/2</u>	<u>5702FR 242573</u>	<u>Eppend roff Model 5702</u>	<u>2020</u>	<u>27/05/2020</u>	<u>100-4400 RPM</u>	<u>20.07.2021</u>	<u>19.07.2022</u>	<u>UltraAir Solutions</u>
15.	Accession	<i>Centrifuge</i>	<i>HAD/C B/7</i>	<i>5702FR 842568</i>	<i>Eppend roff Model 5702</i>	<i>2020</i>	<i>27/05/2020</i>	<i>100-4400 RPM</i>	<i>20.07.2021</i>	<i>19.07.2022</i>	<i>UltraAir Solutions</i>

MPP
DR. PARAG BINIWALE
 M.D., (OB-GYN)
 Reg. No. 62623

Issue No. 02	Issue Date: 05.07.2021	Prepared By: <i>Mangesh</i> QM	Copy No.	Page 3 of 23
Rev. No.: 00	Rev. Date: Nil	Approved By: <i>Mangesh</i> <i>Mangesh</i>	Issued By: <i>Mangesh</i> <i>Mangesh</i>	

A. Patankar
Director,
 PATANKAR HOSPITAL PUNE





HEALTH ACCURATE DIAGNOSTICS

Calibration Record

Format No: HAD-LAB/FM/44

Sl. No	Discipline	Name of equipment	UID (unique identification number)	Sr. No	Model/Make	Year of make	Date of Installation	Range and Accuracy	Date of last calibration	Calibration Due on	Calibrated By
16.	Clinical Pathology	Urine Analyser	HAD/C P/1	302940	SIEMENS CLINITEK STATU S+	2020	16-03-2020	As mentioned in the scope of accreditation			Not Applicable
17.	Cytology	Cyto Centrifuge	HAD/C YT/1	OD000 409	Centrifuge	2017	04-08-2020	2500 RPM	20.07.2021	19.07.2022	Ultra AI Solution Eppendorf
18.	Hematology	Micropipette	HAD/P IP/35	QW028 53	Agappe (Mispa-I Series)/ Agappe	2021	15-07-2021	20-200 µL	15.07.2021	15.07.2022	Eppendorf
19.	Hematology	Micropipette	HAD/P IP/10	I16722 G	Eppendorf (Research Plus)/ Eppendorf	2021	15-07-2021	2-20 µl	15.07.2021	15.07.2022	Eppendorf
20.	Hematology	Micropipette	HAD/P IP/13	JW063 23	Thermo Scientific	2021	15-07-2021	100-1000 µl	15.07.2021	15.07.2022	Eppendorf

Signature

Director,

Issue No. 02	Issue Date: 05.07.2021	Prepared By: <i>Signature</i> QM	Copy No.
Rev. No. 00	Rev. Date: Nil	Approved By: <i>Signature</i>	Issued By: <i>Signature</i>

DR. PARAG BINIWA
M.D., (OB-G)
Reg. No. 62623

HEALTH ACCURATE DIAGNOSTICS

Calibration Record

Format No: HAD-LAB/FM/44

Sl. No	Discipline	Name of equipment	UID (unique identification number)	Sr. No	Model/M ake	year of make	Date of Installation	Range and Accuracy	Date of last calibration	Calibration Due on	Calibrated By
--------	------------	-------------------	---	--------	----------------	-----------------	-------------------------	-----------------------	-----------------------------	-----------------------	---------------

16.	Clinical Pathology	Urine Analyser	HAD/C P/1	302940	SIEME NS CLINI TEK STATU S+	2020	16-03-2020	As mentioned in the scope of accreditation	Not Applicable		
17.	Cytology	Cyto Centrifuge	HAD/C YT/1	OD000 409	Centris ure	2017	04-08-2020	2500 RPM	20.07.2021	19.07.2022	Ultra Air Solutions
18.	Hematology	Micropipette	HAD/P IP/35	QW028 53	Agappe (Mispa- I Series)/ Agappe	2021	15-07-2021	20-200 µL	15.07.2021	15.07.2022	Eppendorf
19.	Hematology	Micropipette	HAD/P IP/10	I16722 G	Eppend rof (Resear ch Plus)/ Eppend rof	2021	15-07-2021	2-20 µl	15.07.2021	15.07.2022	Eppendorf
20.	Hematology	Micropipette	HAD/P IP/13	JW063 23	Thermo Scientif	2021	15-07-2021	100-1000 µl	15.07.2021	15.07.2022	Eppendorf

Issue No. 02	Issue Date: 05.07.2021	Prepared By: <i>Manglik</i> QM	Copy No.	Page 4 of 23
Rev. No. 00	Rev. Date: Nil	Approved By: <i>Manglik</i> <i>Manglik</i>	Issued By: <i>Manglik</i> <i>Manglik</i>	

A. Patankar

Director,
PVT. LTD.
PVT. LTD.
PVT. LTD.

Manglik
DR. PARAG BINIWALE
M.D., (OB-GYN)
Reg. No. 62623

HEALTH ACCURATE DIAGNOSTICS

Calibration Record

Format No: HAD-LAB/FM/44

Sl. No	Discipline	Name of equipment	UID (unique identification number)	Sr. No	Model/Make	year of make	Date of Installation	Range and Accuracy	Date of last calibration	Calibration Due on	Calibrated By
21.	Hematology	Micropipette	HAD/IP/14	KW04370	Thermo Scientific (Finnpipette F3)/ Thermo Scientific	2021	15-07-2021	5-50 µl	15.07.2021	15.07.2022	Eppendorf
22.	Clinical Biochemistry	Micropipette	HAD/IP/25	Q55670 I	Eppendorf Research plus	2020	08-08-2020	0.5-5ml	11/05/2022	11/05/2023	Eppendorf
23.	Hematology	Haematology Analyser	HAD/H/12	20390	SYSM EX XN 550/SY	2021	08-03-2021	As Defined in the scope	07/03/2022	06/03/2023	Sysmex

Issue No. 02	Issue Date: 05.07.2021	Prepared By: <i>Mangesh</i> QM	Copy No.	Page 5 of 23
Rev. No.: 00	Rev. Date: Nil	Approved By: <i>[Signature]</i>	Issued By: <i>Mangesh</i> (<i>Mangesh</i>)	

Pataunkar
Director,
PATANKAR HOSPITAL PVT. L.



DR. PARAG BINWALE
 M.D. (OB-GYN)
 Reg. No. 62628
 15/7/21



HEALTH ACCURATE DIAGNOSTICS

Calibration Record

Format No: HAD-LAB/FM/44

Sl. No	Discipline	Name of equipment	UID (unique identificati on number)	Sr. No	Model/M ake	year of make	Date of Installation	Range and Accuracy	Date of last calibration	Calibration Due on	Calibrated By
24.	Hematology	ID CENTRIFUGE 12SII	HAD/H /3	-	SMEX BIORA D Diamed /BIOR AD	2014	01-10-2014	1030 RPM	20.07.2021	19.07.2022	Ultra Air Solutions
25.	Microbiology	Biosafety Cabinet	HAD/B SC/03 QC	-	UltraAi r Solutio ns			NA	09/02/2022	08/02/2023	UltraAir Solution
26.	Microbiology	Incubator	HAD/I NC/02	574	DSCO8 /Medic a Instrum ent	2020	06-01-2020	Upto 70°C	20.07.2021	19.07.2022	Ultra A Solutio
27.	Hematology	Flowcytometer	HAD/F C/1	R65687 459021 4	BD FACSV iaTM system/ BD	2019	10-04-2019	As defined in the scope	NA	NA	BD
28.	Molecular biology	Automated Nucleic Acid Extraction	HAD/ MB/17	711- 86070	Thermo Scientif	2020	16-12-2020	NA		NA	N

Issue No. 02	Issue Date: 05.07.2021	Prepared By: <i>Mangela</i> QM	Copy No.	Page 6 of 23
Rev. No.: 00	Rev. Date: Nil	Approved By: <i>Mangela</i>	Issued By: <i>Mangela</i>	

AD Patankar

Mangela
DR. PARAG BINW
M.D., (OB-
Reg. No. 62623

HEALTH ACCURATE DIAGNOSTICS

Calibration Record

Format No: HAD-LAB/FM/44


Sl. No	Discipline	Name of equipment	UID (unique identification number)	Sr. No	Model/Make	year of make	Date of Installation	Range and Accuracy	Date of last calibration	Calibration Due on	Calibrated By
--------	------------	-------------------	---	--------	------------	-----------------	-------------------------	-----------------------	-----------------------------	-----------------------	---------------

		System			ic Kingfisher Flex/Thermo Scientific						
29.	Molecular biology	Automated Nucleic Acid Extraction System	HAD/MB/18	711-8G0781	Thermo Scientific Kingfisher Flex/Thermo Scientific	2021	05-02-2021	NA	NA	NA	NA
30.	Molecular biology	Automated Nucleic Acid Extraction System	HAD/MB/16	EZYE XM300 020090 0961	ZYBIO /EXM3 000/ Zybio	2020	01-10-2020	NA	30.09.2021	30.09.2022	Zybio
31.	Molecular biology	Micropipette	HAD/IP/28	HH513 23	Finnpipette F2/	2021	05-02-2021	2-20 µL	11/05/2022	11/05/2023	Eppendorf

Issue No. 02	Issue Date: 05.07.2021	Prepared By: <i>Mangesh</i> QM	Copy No.	Page 7 of 23
Rev. No.: 00	Rev. Date: Nil	Approved By: <i>Mangesh</i>	Issued By: <i>Mangesh</i>	

AP Patankar

Director,
PATANKAR HOSPITAL PVT. LTD.



HEALTH ACCURATE DIAGNOSTICS

Calibration Record

Format No: HAD-LAB/FM/44

Sl. No	Discipline	Name of equipment	UID (unique identification number)	Sr. No	Model/M ake	year of make	Date of Installation	Range and Accuracy	Date of last calibration	Calibration Due on	Calibrated By
					Thermo Scientif ic						
32.	Molecular biology	Micropipette	HAD/P IP/15	O4370 61	Eppend orf Researc h plus/ Eppend orf	2021	05-02-2021	0.5-10 µL	11/05/2022	11/05/2023	Eppendorf
33.	Molecular biology	Micropipette Mulitchannel	HAD/P IP/27	P64990 I	Eppend orf Researc h plus/ Eppend orf	2021	21-07-2021	0.5-10 µL	20.07.2021	19.07.2022	Eppendorf
34.	Molecular biology	Micropipette	HAD/P IP/16	O4370 7I	Eppend orf Researc h plus/ Eppend orf	2021	05-02-2021	0.5-10 µL	11/05/2022	11/05/2023	Eppendorf

Patankar

**Director,
PATANKAR HOSPITAL PVT. LTD.**

Issue No. 02	Issue Date: 05.07.2021	Prepared By: <i>Mangesh</i> QM	Copy No.	Page 8 of 23
Rev. No.: 00	Rev. Date: Nil	Approved By: <i>[Signatures]</i>	Issued By: <i>Mangesh</i> (<i>Mangesh</i>)	

patankar
DR. PARAG BINIWALE
M.D., (OB-GYN)
Reg. No. 62623

HEALTH ACCURATE DIAGNOSTICS

Calibration Record

Format No: HAD-LAB/FM/44

Sl. No	Discipline	Name of equipment	UID (unique identification number)	Sr. No	Model/Make	year of make	Date of Installation	Range and Accuracy	Date of last calibration	Calibration Due on	Calibrated By
35.	Molecular biology	Micropipette	HAD/P IP/04	KW04 365	Fine pipette F3/ Thermo Scientif ic	2020	04-08-2020	5-50 µL	11/05/2022	11/05/2023	Eppendorf
36.	Molecular biology	Micropipette	HAD/P IP/21	O3391 4I	Eppend orf Researc h plus/ Eppend orf	2021	05-02-2021	100-1000 µL	11/05/2022	11/05/2023	Eppendorf
37.	Molecular biology	Micropipette Mulitchannel	HAD/P IP/29	K4283 8J	Eppend orf Researc h plus/ Eppend orf	2021	16-07-2021	0.5-10 µL	16.07.2021	15.07.2022	Eppendorf
38.	Molecular biology	Micropipette	HAD/P IP/23	O1631 7I	Eppend orf Researc h plus/	2021	05-02-2021	20-200 µL	11/05/2022	11/05/2023	Eppendorf

Issue No. 02	Issue Date: 05.07.2021	Prepared By: <i>Mangesh</i> QM	Copy No.	Page 9 of 23
Rev. No.: 00	Rev. Date: Nil	Approved By: <i>Mangesh</i>	Issued By: <i>Mangesh</i>	

MPG
DR. PARAG BINIWALE
M.D., (OB-GYN)
Reg. No. 62623

AP Patankar
Director,
PATANKAR HOSPITAL PVT. LTD.



HEALTH ACCURATE DIAGNOSTICS

Calibration Record

Format No: HAD-LAB/FM/44

Sl. No	Discipline	Name of equipment	UID (unique identification number)	Sr. No	Model/M ake	year of make	Date of Installation	Range and Accuracy	Date of last calibration	Calibration Due on	Calibrated By
39.	Molecular biology	Micropipette	HAD/P IP/20	L20950 I	Eppend orf Researc h plus/Ep pendorf	2021	05-02-2021	2-20 µL	11/05/2022	11/05/2023	Eppendorf
40.	Molecular biology	<u>Micropipette</u>	<u>HAD/P IP/19</u>	<u>L20674 I</u>	<u>Eppend orf Researc h plus</u>	<u>2021</u>	<u>05-02-2021</u>	<u>2-20 µL</u>	<u>11/05/2022</u>	<u>11/05/2023</u>	<u>Eppendorf</u>
41.	Molecular biology	<u>Micropipette Multichannel</u>	<u>HAD/P IP/30</u>	<u>I18208 J</u>	<u>Eppend orf Researc h plus</u>	<u>2021</u>	<u>16-07-2021</u>	<u>10-100 µL</u>	<u>16.07.21</u>	<u>16.07.22</u>	<u>Eppendorf</u>
42.	Molecular biology	<u>Micropipette Multichannel</u>	<u>HAD/P IP/31</u>	<u>K3791 1J</u>	<u>Eppend orf Researc h plus</u>	<u>2021</u>	<u>21-07-2021</u>	<u>30-300 µL</u>	<u>21.07.21</u>	<u>21.07.22</u>	<u>Eppendorf</u>

Issue No. 02	Issue Date: 05.07.2021	Prepared By: <i>Mangesh</i> QM	Copy No.	Page 10 of 23
Rev. No.: 00	Rev. Date: Nil	Approved By: <i>Mangesh</i>	Issued By: <i>Mangesh</i>	

Patankar
Director,
PATANKAR HOSPITAL PVT. LTD.

MPP
DR. PARAG BINWALE
M.D., (OB-GYN)
Reg. No. 62623

HEALTH ACCURATE DIAGNOSTICS

Calibration Record

Format No: HAD-LAB/FM/44


Sl. No	Discipline	Name of equipment	UID (unique identification number)	Sr. No	Model/M ake	year of make	Date of Installation	Range and Accuracy	Date of last calibration	Calibration Due on	Calibrated By
43.	Molecular biology	<u>Micropipette</u>	<u>HAD/P IP/24</u>	<u>Q1642 4I</u>	<u>Eppend orf Researc h plus</u>	<u>2021</u>	<u>05-02-2021</u>	<u>20-200 µL</u>	<u>21/07/2021</u>	<u>21/07/2022</u>	<u>Eppendorf</u>
44.	Molecular biology	<u>Micropipette</u>	<u>HAD/P IP/32</u>	<u>Q2738 9J</u>	<u>Eppend orf Researc h plus</u>	<u>2021</u>	<u>22-07-2021</u>	<u>100-1000 µL</u>	<u>22.07.2021</u>	<u>22.07.2022</u>	<u>Eppendorf</u>
45.	Molecular biology	<u>Micropipette</u>	<u>HAD/P IP/33</u>	<u>Q2668 5J</u>	<u>Eppend orf Researc h plus</u>	<u>2021</u>	<u>21-07-2021</u>	<u>100-1000 µL</u>	<u>21.07.2021</u>	<u>21.07.2022</u>	<u>Eppendorf</u>
46.	Molecular biology	Thermomixer F 1.5	HAD/ MB/4	5384CI 000567	Eppend orf thermo mixer F 1.5/ 5384 Hambu rg	2014	21-01-2014	Amb to 100°C	06.08.2021	05.08.2022	UltraAirSolu tions

Parag Biniwale

Director,
ANKAR HOSPITAL PVT. LTD.

Issue No. 02	Issue Date: 05.07.2021	Prepared By: <i>Mangesh</i> QM	Copy No.	Page 11 of 23
Rev. No.: 00	Rev. Date: Nil	Approved By: <i>Mangesh</i>	Issued By: <i>Mangesh</i>	

Parag Biniwale
DR. PARAG BINIWALE
M.D., (OB-GYN)
Reg. No. 62623





health
accurate
diagnostics

HEALTH ACCURATE DIAGNOSTICS

Calibration Record

Format No: HAD-LAB/FM/44

Sl. No	Discipline	Name of equipment	UID (unique identification number)	Sr. No	Model/M ake	year of make	Date of Installation	Range and Accuracy	Date of last calibration	Calibration Due on	Calibrated By
47.	Molecular biology	Eppendorf centrifuge 5420 (To be done)	HAD/ MB/5	5420JG 802987	5420/E ppendo rf	2020	06-08-2020	UPTO 15060 RPM	04.08.2021	03.08.2022	Eppendorf
48.	Molecular biology	Deep freezer	HAD/ MB/6	200414 86	NFG45 0/cell frost	2020	06-08-2020	-18 to -22°C	20.07.2021	19.07.2022	UltraAirSolu tions
49.	Molecular biology	Deep freezer	HAD/ MB/7	703531 3	NFG20 0/cell frost	2020	06-08-2020	-20 °C	20.07.2021	19.07.2022	UltraAirSolu tions
50.	Molecular biology	Deep Freezer	HAD/ MB/8	703531 7	Celfrost NFG20 0	2018	10-08-2018	-18 to -22°C	20.07.2021	19.07.2022	UltraAir Solutions
51.	Biochemistry	Deep Freezer	HAD/I A/10	131139 469	Celfrost BFS- 150	2013	22-11-2013	-10 to -40°C	20.07.2021	19.07.2022	UltraAir Solutions
52.	Biochemistry	Refrigerator	HAD/I A/7	130622 9862	Celfrost FKG- 400	2013	6-6-2013	2-8°C	20.07.2021	19.07.2022	UltraAir Solutions
53.	Hematology + Serology	Refrigerator	HAD/C B/8	M1446 102993 900004 1	Trufros † VC-440	2020	27/05/2020	2-8°C	20.07.2021	19.07.2022	UltraAir Solutions

Issue No. 02	Issue Date: 05.07.2021	Prepared By: <i>Mangesh</i> QM	Copy No.	Page 12 of 23
Rev. No.: 00	Rev. Date: Nil	Approved By: <i>Mangesh</i> <i>Mangesh</i>	Issued By: <i>Mangesh</i> <i>Mangesh</i>	

AP Patil

Director,
HOSPITAL PVT. LTD.

MPL
DR. PARAG BINIWALE
M.D., (OB-GYN)
Reg. No. 62623

HEALTH ACCURATE DIAGNOSTICS

Calibration Record

Format No: HAD-LAB/FM/44

Sl. No	Discipline	Name of equipment	UID (unique identification number)	Sr. No	Model/Make	year of make	Date of Installation	Range and Accuracy	Date of last calibration	Calibration Due on	Calibrated By
54.	Biochemistry	Refrigerator	HAD/C B/9	130629 867	Celfrost FKG-400	2013	6-6-2013	2-8°C	20/07/2021	19/07/2022	UltraAir Solutions
55.	Microbiology	Refrigerator	HAD/H /9	FKG32 016133 343	Celfrost FKG-320	2020	27/05/2020	2-8°C	20.07.2021	19.07.2022	UltraAir Solutions
56.	<u>Microbiology & Hematology</u>	<u>Refrigerator</u>	<u>HAD/R EF/01</u>	<u>904NY LX001 538</u>	<u>LG</u>	<u>2020</u>	<u>27/05/2020</u>	<u>2-8oC -15-30 oC</u>	<u>20.07.2021</u>	<u>19.07.2022</u>	<u>UltraAir Solutions</u>
57.	Biochemistry	Refrigerator	HAD/R EF/02		VOLT AS			2-8°C	09/02/2022	08/02/2023	UltraAir Solutions
58.	Hematology	Refrigerator	HAD/R EF/03		VOLT AS			2-8°C	09/02/2022	08/02/2023	UltraAir Solutions
59.	Pantry	Refrigerator	HAD/R EF/04		VOLT AS			2-8°C	09/02/2022	08/02/2023	UltraAir Solutions

Issue No. 02	Issue Date: 05.07.2021	Prepared By: <i>Mangal</i> QM	Copy No.	Page 13 of 23
Rev. No.: 00	Rev. Date: Nil	Approved By: <i>Mangal</i>	Issued By: <i>Mangal</i> (<i>Mangal</i>)	

A. Patankar
Director,
TANKAR HOSPITAL PVT. LTD.

Mangal
DR. PARAG BINWALE
M.D., (OB-GYN)
Reg. No. 62623



HEALTH ACCURATE DIAGNOSTICS

Calibration Record

Format No: HAD-LAB/FM/44

Sl. No	Discipline	Name of equipment	UID (unique identificati on number)	Sr. No	Model/M ake	year of make	Date of Installation	Range and Accuracy	Date of last calibration	Calibration Due on	Calibrated By
60.	Molecular Biology	Biorad Real Time Thermal Cycler CFX96	HAD/ MB/9	787BR 11140	C1000 / BIORA D	2020	04-08-2020	As per the applied scope		NA	
61.	Molecular Biology	Biorad Real Time Thermal Cycler CFX96 (To be added)	HAD/ MB/9A	785BR 23408	CFX 96 TOUC H /BIOR AD	2021	23-11-2021	As per the applied scope		NA	
62.	Molecular Biology	Biorad Real Time Thermal Cycler CFX96 (To be added)	HAD/ MB/9B	785BR 29889	CFX 96 TOUC H /BIOR AD	2021	17.05.2021	As per the applied scope		NA	
63.	Molecular Biology	UV-C Disinfection System (To be removed)	HAD/ MB/10	NA	UASPL Model UAS- UV-41	2020	01-08-2020	NA	20.07.2021	19.07.2022	UltraAirSolu tions
64.	Microbiology	Incubator	HAD/ M/1	10343	BTI-25 / BIOTE CHNIC	2018	12-02-2019	37.55°C	09/02/2022	08/02/2023	UltraAirSolu tions

Issue No. 02	Issue Date: 05.07.2021	Prepared By: <i>Mangesh</i> QM	Copy No.	Page 14 of 23
Rev. No.: 00	Rev. Date: Nil	Approved By: <i>Mic</i> <i>Mic</i>	Issued By: <i>Mangesh</i> (<i>Mangesh</i>)	

HEALTH ACCURATE DIAGNOSTICS

Calibration Record

Format No: HAD-LAB/FM/44

Sl. No	Discipline	Name of equipment	UID (unique identification number)	Sr. No	Model/M ake	year of make	Date of Installation	Range and Accuracy	Date of last calibration	Calibration Due on	Calibrated By
					S INDIA						
65.	Molecular Biology	Autoclave	HAD/G EN/ 06	N/F	SEMIA UTOM ATIC / N/F	2020	20-07-2020	upto 200 °C, 0 to 2.1 kg/cm ² , 0 to 60 min	20.07.2021	19.07.2022	UltraAirSolu tions
66.	Microbiology	Autoclave	HAD/A UT/01	EN831 -1	ANAL OG Model- SS316			upto 200 °C, 0 to 2.1 kg/cm ² , 0 to 60 min	09/02/2022	08/02/2023	UltraAir Solutions
67.	ACCESSION	Micropipette	HAD/P IP/01	870304 4	Proline Plus / Biohit	2017	23-05-2017	20-200 µL	11/05/2022	11/05/2023	Eppendorf
68.	ACCESSION	Micropipette	HAD/P IP/09	I17159 G	Eppend rof (Resear ch Plus)	2021	21-07-2021	100-1000 µL	21.07.2021	21.07.2022	Eppendorf
69.	Molecular Biology	Micropipette	HAD/P IP/02	870275 7	Proline Plus / Biohit	2017	23-05-2017	100-1000 µL	15.07.2021	15.07.2022	Eppendorf

P. Patankar
DIRECTOR,
PANKAR HOSPITAL PVT. LTD.

Issue No. 02	Issue Date: 05.07.2021	Prepared By: <i>Mangesh</i> QM	Copy No.	Page 15 of 23
Rev. No.: 00	Rev. Date: Nil	Approved By: <i>Mangesh</i> <i>Mangesh</i>	Issued By: <i>Mangesh</i> (<i>Mangesh</i>)	

MPL
DR. PARAG BINIWALE
M.D., (OB-GYN)
Reg. No. 62623



HEALTH ACCURATE DIAGNOSTICS

Calibration Record

Format No: HAD-LAB/FM/44

Sl. No	Discipline	Name of equipment	UID (unique identificati on number)	Sr. No	Model/M ake	year of make	Date of Installation	Range and Accuracy	Date of last calibration	Calibration Due on	Calibrated By
70.	Molecular Biology	Micropipette	HAD/P IP/03	850869 2	Proline Plus / Biohit	2017	23-05-2017	2-20 µL	21.07.2021	21.07.2022	Eppendorf
71.	Molecular Biology	Micropipette	HAD/P IP/08	HH667 74	Thermo Scientif ic (Finnpi pette F2)	2021	15-07-2021	20-200 µl	15.07.2021	15.07.2022	Eppendorf
72.	Molecular Biology	Micropipette	HAD/P IP/07	I51566 G	Eppend orf	2017	23-05-2017	20-200 µL	15.07.2021	15.07.2022	Eppendorf
73.	Microbiology	Thermometer	HAD/C B/12	-	ST- 9283B / Multi	2017	31-05-2017	-50 °C to 300°C	20.07.2021	19.07.2022	UltraAirSolu tions
74.	Hematology	THERMO HYGROMETER	HAD/A CC/1	N/F	THC- 10 / OPTIL AB	2020	27-05-2020	-10 °C to +50°C 20 to 99%RH	20.07.2021	19.07.2022	UltraAirSolu tions
75.	Molecular Biology	THERMO HYGROMETER	HAD/X V/1	N/F	THC- 10 / OPTIL	2020	27-05-2020	-10 °C to +50°C 20 to 99%RH	20.07.2021	19.07.2022	UltraAirSolu tions

Issue No. 02	Issue Date: 05.07.2021	Prepared By: <i>Mangesh</i> QM	Copy No.	Page 16 of 23
Rev. No.: 00	Rev. Date: Nil	Approved By: <i>Mangesh</i> <i>Mangesh</i>	Issued By: <i>Mangesh</i> <i>Mangesh</i>	

AP Patankar
Director,
PATANKAR HOSPITAL PVT. LTD.

MBZ
DR. PARAG BINWALE
M.D., (OB-GYN)
Reg. No. 62623

HEALTH ACCURATE DIAGNOSTICS

Calibration Record

Format No: HAD-LAB/FM/44

Sl. No	Discipline	Name of equipment	UID (unique identification number)	Sr. No	Model/M ake	year of make	Date of Installation	Range and Accuracy	Date of last calibration	Calibration Due on	Calibrated By
					AB						
76.	Molecular Biology	THERMO- HYGROMETER	HAD/ MB/5	-	THC- 10/ OPTIL AB	2020	27/05/2020	-10 °C to +50°C 20 to 99%RH	20.07.2021	19.07.2022	UltraAirSolu tions
77.	Biochemistry	THERMO- HYGROMETER	HAD/P H/2	-	THC- 10/ OPTIL AB	2020	27/05/2020	-10 °C to +50°C 20 to 99%RH	20.07.2021	19.07.2022	UltraAirSolu tions
78.	Clinical Pathology	THERMO- HYGROMETER	HAD/C P/4	-	THC- 10/ OPTIL AB	2020	27/05/2020	-10 °C to +50°C 20 to 99%RH	20.07.2021	19.07.2022	UltraAirSolu tions
79.	Molecular biology	THERMOHYGRO METER CRTIFICATE ABSENT	HAD/D TH/02	212007 03	THC- 10/ OPTIL AB	2020	27/05/2020	-10 °C to +50°C 20 to 99%RH	20/07/2021	19/07/2022	UltraAirSolu tions
80.	Molecular biology	THERMOHYGRO METER CRTIFICATE	HAD/D TH/01	212007 01	THC- 10/ OPTIL	2020	27/05/2020	-10 °C to +50°C 20 to 99%RH	20/07/2021	19/07/2022	UltraAirSolu tions

Issue No. 02	Issue Date: 05.07.2021	Prepared By: <i>Mangesh</i> QM	Copy No.	Page 17 of 23
Rev. No.: 00	Rev. Date: Nil	Approved By: <i>Mangesh</i> <i>Mangesh</i>	Issued By: <i>Mangesh</i> <i>Mangesh</i>	

Mataurkar
Director,
PATAANKAR HOSPITAL PVT. LTD.

Mangesh
DR. PARAG BINWALE
M.D., (OB-GYN)
Reg. No. 62623



HEALTH ACCURATE DIAGNOSTICS

Calibration Record

Format No: HAD-LAB/FM/44

Sl. No	Discipline	Name of equipment	UID (unique identification number)	Sr. No	Model/Make	year of make	Date of Installation	Range and Accuracy	Date of last calibration	Calibration Due on	Calibrated By
		ABSENT			AB						
81.	Microbiology	THERMOHYGROMETER CERTIFICATE ABSENT	HAD/DTH/03	21200704	THC-10/OPTILAB	2020	27/05/2020	-10 °C to +50°C 20 to 99%RH	20/07/2021	19/07/2022	UltraAirSolutions
82.	Hematology	Digital Thermometer	HAD/DTM/01	-	Multi/ST-9283B	2020	May 2020	-50 °C to 300°C	20.07.2021	19.07.2022	UltraAir Solutions
83.	Hematology	Digital Thermometer	HAD/DTM/04	-	Multi/ST-9283B	2020	May 2020	-50 °C to 300°C	20.07.2021	19.07.2022	UltraAir Solutions
84.	Hematology	Digital Thermometer	HAD/GEN/3	-	Multi/ST-9283B	2020	May 2020	-50 °C to 300°C	20.07.2021	19.07.2022	UltraAir Solutions
85.	Clinical Biochemistry	DIGITAL THERMOMETER	HAD/DTM/02	210517	MULTI / ST-9283B	2021	17/05/2021	-50 °C to +300°C	20.07.2021	19.07.2022	UltraAirSolutions

Patankar

Director,
PATANKAR HOSPITAL PVT. LTD.

Issue No. 02	Issue Date: 05.07.2021	Prepared By: <i>Mangesh</i> QM	Copy No.	Page 18 of 23
Rev. No.: 00	Rev. Date: Nil	Approved By: <i>Mic Mic</i>	Issued By: <i>Mangesh</i> (<i>Mangesh</i>)	

pppl

DR. PARAG BINIWALE
M.D., (OB-GYN)
Reg. No. 62623

HEALTH ACCURATE DIAGNOSTICS

Calibration Record

Format No: HAD-LAB/FM/44

Sl. No	Discipline	Name of equipment	UID (unique identificati on number)	Sr. No	Model/M ake	year of make	Date of Installation	Range and Accuracy	Date of last calibration	Calibration Due on	Calibrated By
86.	Clinical Biochemistry	DIGITAL THERMOMETER	HAD/D TM/03	210518	MULTI / ST- 9283B	2021	17/05/2021	-50 °C to +300°C	20.07.2021	19.07.2022	UltraAirSolu tions
87.	Accession	GLASS THERMOMETER	HAD/G T/01	212007 07	JRM	2021	20/07/2021	-50 °C to +50°C	20.07.2021	19.07.2022	UltraAirSolu tions
88.	Accession	GLASS THERMOMETER	HAD/G T/02	212007 08	JRM	2021	20/07/2021	-50 °C to +50°C	20.07.2021	19.07.2022	UltraAirSolu tions
89.	Accession	Glass Thermometer	HAD/G T/03	212007 09	JRM	2021	20/07/2021	-50 °C to +50°C	20.07.2021	19.07.2022	UltraAirSolu tions
90.	Accession	Glass Thermometer	HAD/G T/04	212007 10	JRM	2021	20/07/2021	-50 °C to +50°C	20.07.2021	19.07.2022	UltraAirSolu tions
91.	Molecular Biology	PCR HOOD	HAD/ MB/13	N/F	Ultra- Air Solutio ns	2020	11-08-2020	NA	21.07.2021	19.07.2022	UltraAirSolu tions

Issue No. 02	Issue Date: 05.07.2021	Prepared By: <i>Mangesh</i> QM	Copy No.	Page 19 of 23
Rev. No.: 00	Rev. Date: Nil	Approved By: <i>Mangesh</i> <i>Mangesh</i>	Issued By: <i>Mangesh</i> <i>Mangesh</i>	

Director,
TANKAR HOSPITAL PVT. LTD.

MPPZ
DR. PARAG BINIWALE
 M.D., (OB-GYN)
 Reg. No. 62623



HEALTH ACCURATE DIAGNOSTICS

Calibration Record

Format No: HAD-LAB/FM/44

Sl. No	Discipline	Name of equipment	UID (unique identification number)	Sr. No	Model/M ake	year of make	Date of Installation	Range and Accuracy	Date of last calibration	Calibration Due on	Calibrated By
92.	Molecular Biology	PCR Hood	HAD/ MB/14	N/F	Ultra- Air Solutio ns	2020	11-08-2020	NA	21.07.2021	19.07.2022	UltraAirSolu tions
93.	Molecular Biology	PCR Plate centrifuge	HAD/ MB/15	040U2 072804 66	BR- PCR- 2200 / BR Bioche m	2020	12-08-2020	2500 rpm	20.07.2021	19.07.2022	UltraAirSolu tions
94.	Molecular Biology	PCR Plate centrifuge	HAD/ MB/19	040U2 081105 21	BR Bioche m BR- PCR- 2200	2021	01/01/2021	2500 RPM	20.07.2021	19.07.2022	UltraAirSolu tions
95.	Hematology	Timer	HAD/D T/02	212007 05	Winner Loudsta r/ 10S	2020	27/05/2020	1Sec to 99 Hrs	20.07.2021	19.07.2022	UltraAirSolu tions
96.	Microbiology	Timer	HAD/D T/03	212007 06	Winner Loudsta r/ 10S	2020	27/05/2020	1Sec to 99 Hrs	20.07.2021	19.07.2022	UltraAirSolu tions

Issue No. 02	Issue Date: 05.07.2021	Prepared By: <i>Mangish</i> QM	Copy No.	Page 20 of 23
Rev. No.: 00	Rev. Date: Nil	Approved By: <i>Mangish</i>	Issued By: <i>Mangish</i>	

Atankar
Director,
ATANKAR HOSPITAL PVT. LTD.

10/7/22
DR. PARAG BINWALE
M.D., (OB-GYN)
Reg. No. 62623

HEALTH ACCURATE DIAGNOSTICS

Calibration Record

Format No: HAD-LAB/FM/44

Sl. No	Discipline	Name of equipment	UID (unique identification number)	Sr. No	Model/Make	year of make	Date of Installation	Range and Accuracy	Date of last calibration	Calibration Due on	Calibrated By
97.	Molecular Biology	Ultra Low Freezer	HAD/R /1	815459 -558	FORM A 900 SERIE S / Model 991 / Thermo SCIEN TIFIC	2018	10.08.2018	-50 to -80 °C	20.07.2021	19.07.2022	UltraAir Solutions
98.	Molecular Biology	Ultra Low Freezer	HAD/R /2	129011 401100 415	REVC O VALU E PLUS / Model No. ULT17 86-4- V49 / Thermo SCIEN TIFIC	2018	10-08-2018	-80 °C	20.07.2021	19.07.2022	UltraAir Solutions

Issue No. 02	Issue Date: 05.07.2021	Prepared By: <i>Manglik</i> QM	Copy No.	Page 21 of 23
No.: 00	Rev. Date: Nil	Approved By: <i>Manglik</i>	Issued By: <i>Manglik</i>	

A. Patil
Director,
ANKAR HOSPITAL PVT. LTD.

Manglik
DR. PARAG BINIWALE
M.D., (OB-GYN)



HEALTH ACCURATE DIAGNOSTICS

Calibration Record

Format No: HAD-LAB/FM/44

Sl. No	Discipline	Name of equipment	UID (unique identificati on number)	Sr. No	Model/M ake	year of make	Date of Installation	Range and Accuracy	Date of last calibration	Calibration Due on	Calibrated By
99.	Molecular Biology	GeneXpert Dx	HAD/ MB/1	821607	Dx / Cephei d	2018	17-06-2019	40 - 10000000 copies/mL	18/12/2021	18/12/2022	Cepheid
100.	Molecular Biology	Mini Centrifuge	HAD/ MB/11	131007 52 C1413 V- 230EU	MiniSta r/ VWR Internat ional	2020	04-08-2020	2000- 3000 rpm	20.07.2021	19.07.2022	UltraAirSolu tions
101.	Molecular Biology	Mini Centrifuge	HAD/ MB/12	HSA17 854	SPROU T/ HEAT HROW	2017	04-08-2020	2000-3000 rpm	20.07.2021	19.07.2022	UltraAirSolu tions
102.	Clinical Pathology	Microscope	HAD/ MIC/1		Olymp us CH20i	2004	FEB 2004	2, 4, 10, 40 X	NA	NA	Biomedical - Engineer Services (Mr Saran)
103.	Microbiology	Microscope	HAD/ MIC/2	026469	LABO MED	2004	FEB 2004	4, 10, 40, 100 X	NA	NA	Biomedical Engineer Services (Mr Saran)

Issue No. 02	Issue Date: 05.07.2021	Prepared By: <i>Mangesh</i> QM	Copy No.	Page 22 of 23
Rev. No.: 00	Rev. Date: Nil	Approved By: <i>Mangesh</i> <i>Mangesh</i>	Issued By: <i>Mangesh</i> (<i>Mangesh</i>)	

AP Patil
Director,
ATANKAR HOSPITAL PVT. LTD.

Mangesh
DR. PARAG BINIWALE
M.D., (OB-GYN)
Reg. No. 62623

HEALTH ACCURATE DIAGNOSTICS

Calibration Record

Format No: HAD-LAB/FM/44

Sl. No	Discipline	Name of equipment	UID (unique identification number)	Sr. No	Model/M ake	year of make	Date of Installation	Range and Accuracy	Date of last calibration	Calibration Due on	Calibrated By
104.	Hematology	Microscope	HAD/ MIC/3	-	Olymp us CH20i	2004	FEB 2004	10, 20, 40, 100 X	NA	NA	Biomedical Engineer Services (Mr Saran)



Director,
PATANKAR HOSPITAL PVT. LTD.



DR. PARAG BINWALE
M.D., (OB-GYN)
Reg. No. 62623

Issue No. 02	Issue Date: 05.07.2021	Prepared By:  QM	Copy No.	Page 23 of 23
Rev. No.: 00	Rev. Date: Nil	Approved By:  	Issued By:  	

